

Ti_{1-x}Au_x Alloys: Hard Biocompatible Metals and Their Possible Applications

E. Svanidze,¹ T. Besara,² M. F. Ozaydin,³ Yan Xin,² K. Han,² H. Liang,³ T. Siegrist,² and E. Morosan¹

¹Department of Physics and Astronomy, Rice University, Houston, TX 77005, USA

²National High Magnetic Field Laboratory, Florida State University, Tallahassee, FL 32306, USA and

³Department of Mechanical Engineering, Texas A&M University, College Station, TX 77843, USA

(Dated: July 9, 2018)

The search for new hard materials is often challenging from both theoretical and experimental points of view. Furthermore, using materials for biomedical applications calls for alloys with high biocompatibility which are even more sparse. The Ti_{1-x}Au_x (0.22 ≤ x ≤ 0.8) exhibit extreme hardness and strength values, elevated melting temperatures (compared to those of constituent elements), reduced density compared to Au, high malleability, bulk metallicity, high biocompatibility, low wear, reduced friction, potentially high radio opacity, as well as osseointegration. All these properties render the Ti_{1-x}Au_x alloys particularly useful for orthopedic, dental, and prosthetic applications, where they could be used as both permanent and temporary components. Additionally, the ability of Ti_{1-x}Au_x alloys to adhere to ceramic parts could reduce the weight and cost of these components.

I. INTRODUCTION

In addition to numerous applications in the industrial, automotive and aerospace fields, Ti has been widely used for implant devices that replace patients' hard tissues [1, 2]. A number of *in vivo* and *in vitro* experiments with various grades of Ti concluded that commercially pure Ti is a highly biocompatible material due to the spontaneous build-up of an inert and stable oxide layer [1, 3]. Additional properties that make Ti suitable for biomedical applications include high strength-to-weight ratio [4, 5], low electrical conductivity, low ion-formation levels in aqueous environments, low pH value, and a dielectric constant comparable to that of water [1]. Moreover, Ti is one of a few materials capable of osseointegration – mechanical retention of the implant by the host bone tissue – which stabilizes the implant without any soft tissue layers between the two [6]. These properties enable a wide use of Ti for devices such as artificial knee and hip joints, screws and shunts for fracture fixation, bone plates, pacemakers and cardiac valve prosthesis [7, 8]. Not surprisingly, the dental applications of Ti are just as common, including implants and their components such as inlays, crowns, overdentures, and bridges [1, 9–13].

However, the pure Ti is not strong enough for a number of medical purposes [14, 15], calling for the development of more superior Ti-based alloys [16–20]. Additionally, Ti exhibits poor machineability, which reduces tool life, increases the processing time and is problematic when the elimination of a dental Ti prosthesis is necessary [16, 21]. Both the machineability and hardness can be improved by alloying Ti with another element. After a number of toxic effects were reported in permanent implants [1], the use of V- and Al-containing Ti alloys was discontinued. Among biocompatible elements, the addition of Ag and Cu nearly doubles the hardness, compared to pure Ti [20, 22–24]. Biocompatible Au, which is located in the same group as Ag and Cu, is a β-stabilizer – Ti oc-

curs in two different crystallographic structures (hexagonal *P6₃/mmc* α-Ti and higher hardness cubic *Im $\bar{3}m$* β-Ti). Since both Ag and Cu are α-stabilizers, α-Ti was produced by alloying Ti with these metals. It can be expected that, if it were possible to form the equivalent β-Ti-Au alloys, their hardness may increase. Moreover, the high biocompatibility and corrosion resistance of Au may yield an alloy suitable for biomedical purposes [10], given the wide use of Au and Au-doped implant devices [25, 25–28]. In case the machineability decreases with increased hardness, the relatively low melting temperatures of Ti-Au alloys will allow for the majority of parts to be used in an as-cast form. Additionally, the Ti-Au alloys can adhere to a ceramic surface, making it convenient for a number of biomedical applications, reducing the overall weight and cost of the corresponding parts [25, 29].

In this manuscript we present the hardness, wear rate as well as the microstructural analysis of Ti_{1-x}Au_x (0.22 ≤ x ≤ 0.8) alloys indicating that the x = 0.25 alloy is suitable for a number of biomedical applications, particularly where Ti is already employed. Despite the fact that Au is a β-Ti stabilizer, it has been observed that the x = 0.25 alloy forms two microstructures: αTi and Ti₃Au. A remarkable nearly four-fold increase in hardness, as compared with pure Ti, is registered for the Ti_{0.75}Au_{0.25} alloy. The hardness values exceed those of most biocompatible alloys and engineering metals, as well as most engineering ceramics.

II. EXPERIMENTAL METHODS

Alloys of Ti_{1-x}Au_x were prepared by arc-melting Ti (Cerac, 99.99%) and Au (Cerac, 99.99%) in stoichiometric ratios, with mass losses no more than 0.3%. To ensure homogeneity, the samples were re-melted several times. The Vickers hardness H_V was estimated using a Tukon




FIG. 1: Hardness H_V as a function of composition for $Ti_{1-x}Au_x$ alloys: current study (circles), Au-rich (triangles) [25] and Ti-rich (diamonds) [13] regions, along with elemental data (squares) [32].

2100 microhardness tester, equipped with a Vickers diamond pyramid indenter. The microhardness tests were performed on a polished sample surface of about 3 mm in diameter. Multiple tests were conducted for all samples, using a 300 g load, with a duration of 10 s. Another important quantity is the yield strength σ_f [30], which, for metals, is defined as the stress at which the stress-strain curve for axial load deviates by a strain of 0.2 % from the linear elastic line [30]. At times when stress-strain data are not available, as is the case in $Ti_{1-x}Au_x$ alloys, the value of σ_f can be extracted from the Vickers hardness H_V using $\sigma_f = H_V/3$ [30]. The H_V is the Vickers hardness in units of MPa, resulting in the same units for σ_f . This equation has been empirically shown to be approximately true for metals [31].

III. DATA ANALYSIS AND DISCUSSION

Although both the Ti-rich and Au-rich sides of the Ti-Au phase diagram were previously explored in detail, hardly anything has been reported on the hardness of intermediate compositions. In order to further investigate this system, a series of $Ti_{1-x}Au_x$ alloys was prepared with $0.22 \leq x \leq 0.8$.

Surprisingly, the measurements reveal a non-monotonous change of hardness with x . Even more remarkable is the broad peak in x between 0.22 and 0.35, where the hardness is up to three or four times higher than that of pure Ti. This indicates that the $Ti_{1-x}Au_x$ alloys with $0.22 \leq x \leq 0.35$ are likely better suited for applications where Ti is currently used [1, 7–12]. Of the measured compositions, the $x = 0.25$ alloy exhibits the highest hardness value of 7.81 GPa, even higher than that of




FIG. 2: Mass density ρ (left axis, circles) and melting temperature T_m [33] (right axis, triangles) as a function of composition for $Ti_{1-x}Au_x$ alloys ($0 \leq x \leq 1$).




FIG. 3: Ashby diagram with strength σ_f vs. mass density ρ for different families of materials [30], along with $Ti_{1-x}Au_x$ alloys (current study, green symbols).

pearlitic steels [34, 35] and similar to that of high-carbon steels [36]. As mentioned above, both Ti and Au are biocompatible and have high resistance to *in vivo* corrosion, suggesting that the resulting alloys are suitable for biomedical applications [1, 3, 10].

Since the mass density ρ of Au is lowered by Ti dilution, as shown in Fig. 2, $Ti_{1-x}Au_x$ alloys are even more desirable for biomedical applications which usually employ the denser Au, given the decreased density ρ and increased hardness H_V , yielding stronger and lighter components with increasing x in $Ti_{1-x}Au_x$. Melting the $Ti_{1-x}Au_x$ in Al_2O_3 crucibles resulted in the intermetallic alloy coating the walls of the ceramic container, making the mechanical removal of the coating impossible. This can reduce both the weight and cost of medical components, by us-

ing $\text{Ti}_{1-x}\text{Au}_x$ as a coating for a ceramic part. Moreover, the previously observed radio opacity of $\text{Ti}_{1-x}\text{Au}_x$ ($0.45 \leq x \leq 0.55$) alloys [37, 38] allows for radiographic distinction between implants and live tissue.

The melting temperatures T_m of the $\text{Ti}_{1-x}\text{Au}_x$ alloys, extracted from the Ti-Au phase diagram [33], are plotted in Fig. 2 and summarized in Table I. The values are higher than those of both Ti and Au, which preliminarily classifies $\text{Ti}_{1-x}\text{Au}_x$ alloys as refractory metals [39]. Detailed wear analysis is necessary to prove the refractory nature of the $\text{Ti}_{1-x}\text{Au}_x$ alloys. The potential applications include high-temperature engineering structures, heat sinks, electronics, aerospace, metal cutting and forming, as well as the tooling industry [40].

Preliminary strength estimates, determined from hardness values, place $\text{Ti}_{1-x}\text{Au}_x$ alloys (green symbols, Fig. 3) just above the engineering metals region of the strength-mass density Ashby diagram. While other metallic alloys such as tungsten carbide, high-carbon steels, and nickel alloys show similar hardness values, they are often not desirable for medical applications due to high density and high toxicity.

Given that Ti alloys are frequently heat-treated to improve both hardness and ductility, annealing studies were carried out for the Ti-Au system. However, annealing using two different recipes - for 7 days at 900°C , or for several hours at $0.5T_m$ and $0.3T_m$ (similar to what has been done for other Ti-base alloys [41]) resulted in minimal changes in the hardness compared to the as-cast samples, which might be caused by variation in microstructure homogeneity which can mask the true annealing effects.

IV. CONCLUSIONS

A series of $\text{Ti}_{1-x}\text{Au}_x$ alloys ($0.22 \leq x \leq 0.8$) has been investigated due to their extreme hardness values, elevated melting temperatures (compared to those of constituent elements), reduced density compared to pure Au, bulk metallicity, high biocompatibility and radio opacity. These properties make medical applications especially favorable with examples including replacement parts and components (both permanent and temporary), dental prosthetics and implants. The ability to adhere to ceramic components along with osseointegration are also beneficial, as they are able to reduce component weight and cost. Possible additional applications include circuit wires, hard coatings for tools and other medical equipment, drill head bits, as well as sporting goods.

ACKNOWLEDGEMENTS

The work at Rice was supported by NSF DMR 0847681 (E.M. and E.S.).

- [1] C. N. Elias, J. H. C. Lima, R. Valiev, and M. A. Meyers, "Biomedical applications of titanium and its alloys," *Journal of Metals*, p. 46, 2008.
- [2] M. Semlitsch, "Titanium alloys for hip joint replacements," *Clinical Materials*, vol. 2, p. 1, 1987.
- [3] C. B. Johansson, *On Tissue Reactions to Metal Implants*. PhD Thesis, University of Goteburg, Sweden, 1991.
- [4] J. A. Hautaniemi, H. Hero, and J. T. Juhanoja, "On the bonding of porcelain on titanium," *Journal of Materials Science: Materials in Medicine*, vol. 3, p. 186, 1992.
- [5] R. Wang and A. Fenton, "Titanium for prosthodontic applications: A review of the literature," *Quintessence International*, vol. 27, p. 401, 1996.
- [6] P. I. Branemark, "Osseointegration and its experimental background," *The Journal of Prosthetic Dentistry*, vol. 50, p. 399, 1983.
- [7] G. V. Samsonov, *Handbook of the Physicochemical Properties of the Elements*. Springer, 1968.
- [8] S. K. S. Marya and R. K. Bawari, *Total Hip Replacement Surgery*. JP Medical Ltd., 2010.
- [9] B. Kempf, H. M. Rigelstein, A. Voelcker, and U. Birkholz, "Dental construction element for use in a casting-on process," *European Patent Application*, vol. EP 729739 A2 19960904, 1996.
- [10] B. Kempf, H. M. Rigelstein, T. Baumgaerther, and L. Voelkl, "Manufacture of cast gold-titanium alloy dental or jewelry parts," *European Patent Application*, vol. EP 717118 A2 19960619, 1996.
- [11] P. J. Cascone, "Dental prosthesis method and alloys," *United States Patent and Trademark Office*, vol. US 7892480 B2, 2011.
- [12] N. Vuilleme, "Non-precious dental alloy," *United States Patent and Trademark Office*, vol. US 6613275 B1, 2003.
- [13] M. Takahashi, M. Kikuchi, and O. Okuno, "Mechanical properties and grindability of experimental Ti-Au alloys," *Dental Materials Journal*, vol. 23, p. 203, 2004.
- [14] W. Ho, W. Chen, S. Wu, and H. Hsu, "Structure, mechanical properties, and grindability of dental TiZr alloys," *Journal of Materials Science: Materials in Medicine*, vol. 19, p. 3179, 2008.
- [15] A. Faria, R. Rodrigues, Q. Claro, W. Mattos, and R. Ribeiro, "Wear resistance of experimental titanium alloys for dental applications," *Journal of the Mechanical Behavior of Biomedical Materials*, vol. 4, p. 1873, 2011.
- [16] O. Okuno and H. Hamanaka, "Application of beta titanium alloys in dentistry," *Journal of Dentistry of Japan*, vol. 26, p. 101, 1989.
- [17] E. P. Lautenschlager and P. Monaghan, "Titanium and titanium alloys as dental materials," *International Dental Journal*, vol. 43, p. 245, 1993.
- [18] M. Hattori, K. Hasegawa, M. Yoshinari, E. Kawada, Y. Oda, and T. Okabe, "The effect of fluoride and albumin on corrosion of titanium," *Dental Materials Journal*, vol. 20, p. 16, 2001.
- [19] Y. Takada, H. Nakajima, O. Okuno, and T. Okabe, "Microstructure and corrosion behavior of binary titanium alloys with beta-stabilizing elements," *Dental Materials Journal*, vol. 20, p. 34, 2001.
- [20] M. Takahashi, M. Kikuchi, Y. Takada, and O. Okuno, "Mechanical properties and microstructures of dental

TABLE I: Summary of parameters for as-cast $\text{Ti}_{1-x}\text{Au}_x$ alloys.

x	Melting temperature T_m [33] (°C)	Mass density ρ (g/cm ³)	As-cast hardness H_V (GPa)	As-cast strength σ_f (GPa)	TEM composition	X-ray composition
0	882	4.45	2.62 ± 0.22	0.89 ± 0.07	–	–
0.22	1367	7.14	7.11 ± 0.12	2.42 ± 0.04	$\text{Ti}_3\text{Au} + \alpha\text{Ti}$	Ti_3Au
0.25	1395	7.87	7.81 ± 0.21	2.65 ± 0.07		Ti_3Au
0.33	1310	8.81	6.15 ± 0.44	2.05 ± 0.1	$\text{TiAu} + \text{TiAu}_2$	
0.45	1470	10.39	2.99 ± 0.40	1.02 ± 0.13		
0.50	1495	11.05	2.34 ± 0.24	0.80 ± 0.08	$\beta\text{TiAu} + \text{Ti}$	TiAu
0.60	1385	12.25	2.64 ± 0.24	0.90 ± 0.08	$\text{TiAu} + \text{TiAu}_2 + \text{TiAu}_4$	TiAu_2
0.67	1455	14.09	2.09 ± 0.05	0.71 ± 0.02		TiAu_2
0.75	1400	14.35	2.81 ± 0.30	0.95 ± 0.10		
0.80	1315	14.52	2.83 ± 0.08	0.96 ± 0.03		TiAu_4
1.00	1064	17.65	0.41 ± 0.04	0.14 ± 0.01	–	–

cast Ti-Ag and Ti-Cu alloys,” *Dental Materials*, vol. 21, p. 270, 2002.

- [21] H. Nakajima and T. Okabe, “Titanium in dentistry: development and research in the U.S.A.” *Dental Materials Journal*, vol. 15, p. 77, 1996.
- [22] M. Kikuchi, Y. Takada, S. Kiyosue, M. Yoda, M. Woldu, Z. Cai, O. Okuno, and T. Okabe, “Mechanical properties and microstructures of cast Ti-Cu alloys,” *Dental Materials*, vol. 19, p. 174, 2003.
- [23] M. Kikuchi, Y. Takada, S. Kiyosue, M. Yoda, M. Woldu, O. Okuno, Z. Cai, and T. Okabe, “Grindability of cast Ti-Cu alloys,” *Dental Materials*, vol. 19, p. 375, 2003.
- [24] M. Kikuchi, M. Takahashi, T. Okabe, and O. Okuno, “Grindability of cast Ti-Ag and Ti-Cu alloys,” *Dental Materials*, vol. 22, p. 191, 2003.
- [25] J. Fischer, “Mechanical, thermal, and chemical analyses of the binary system Au-Ti in the development of a dental alloy,” *Journal of Biomedical Materials Research*, vol. 52, p. 678, 2000.
- [26] G. Gafner, “The development of 990 Gold-Titanium: its production, use and properties,” *Gold Bulletin*, vol. 22, p. 112, 1989.
- [27] G. Humpston and D. M. Jacobson, “A new high strength gold bond wire,” *Gold Bulletin*, vol. 25, p. 132, 1992.
- [28] T. Takahashi, M. Kikuchi, Y. Takada, and O. Okuno, “Basic compositions of gold-titanium alloys for dental casting,” *Japanese Society for Dental Materials and Devices*, vol. 17, p. 126, 1998.
- [29] R. Strietzel, “Burning alloy for the production of ceramically veneered dental restorations,” *United States Patent and Trademark Office*, vol. US 7491361 B2, 2009.
- [30] M. F. Ashby, *Materials Selection in Mechanical Design*. Butterworth-Heinemann, 2011.
- [31] D. Tabor, *The Hardness of Metals*. Clarendon Press, Oxford, UK, 1951.
- [32] K. Edalati and Z. Horita, “Universal plot for hardness variation in pure metals processed by high-pressure torsion,” *Materials Transactions*, vol. 51, p. 1051, 2010.
- [33] P. Villars, H. Okamoto, and K. Cenual, *ASM Alloy Phase Diagrams Center*. ASM International, Metals Park, Ohio, 2007.
- [34] K. Han, D. V. Edmonds, and G. D. W. Smith, “Optimization of mechanical properties of high-carbon pearlitic steels with Si and V additions,” *Metallurgical and Materials Transactions A*, vol. 32, p. 1314, 2001.
- [35] D. Raabe, P. Choi, Y. Li, A. Kostka, X. Sauvage, F. Lecouturier, K. Hono, R. Kirchheim, R. Pippan, and D. Embury, “Metallic composites processed via extreme deformation: toward the limits of strength in bulk materials,” *MRS Bulletin*, vol. 35, p. 982, 2010.
- [36] L. Cui, H. Fujii, N. Tsuji, and K. Nogi, “Friction stir welding of a high carbon steel,” *Scripta Materialia*, vol. 56, p. 637, 2007.
- [37] J. F. Boylan, W. Cornish, and J. A. Simpson, “Radiopaque markers and medical devices comprising binary alloys of titanium,” *United States Patent and Trademark Office*, vol. US20090099645 A1, 2009.
- [38] J. F. Boylan and J. A. Simpson, “Radiopaque markers comprising binary alloys of titanium,” *United States Patent and Trademark Office*, vol. US20080288056 A1, 2008.
- [39] C. S. Young and J. C. Durham, *Industrial Applications of Titanium and Zirconium*. ASTM International, West Conshohocken, PA, 1986.
- [40] R. E. Smallwood, *Refractory Metals and their Industrial Applications*. ASTM International, West Conshohocken, PA, 1983.
- [41] T. V. Rajan, C. P. Sharma, and A. Sharma, *Heat Treatment: Principles and Techniques*. PHI Learning Private Limited, New Delhi, 1994.