

Liquid-shaped microlens for scalable production of ultrahigh-resolution OCT microendoscope

Chao Xu¹, Xin Guan², Syeda Aimen Abbasi¹, Neng Xia³, To Ngai², Li Zhang³, Ho-Pui Ho¹, Sze Hang Calvin Ng⁴, and Wu Yuan^{1,*}

¹Department of Biomedical Engineering, The Chinese University of Hong Kong, Hong Kong SAR, China

²Department of Chemistry, The Chinese University of Hong Kong, Hong Kong SAR, China

³Department of Mechanical and Automation Engineering, The Chinese University of Hong Kong, Hong Kong SAR, China

⁴Department of Surgery, The Chinese University of Hong Kong, Hong Kong SAR, China

*wyuan@cuhk.edu.hk

Abstract

Endoscopic optical coherence tomography (OCT) is a valuable tool for providing diagnostic images of internal organs and guiding interventions in real time. Miniaturized OCT endoscopes are essential for imaging small and convoluted luminal organs while minimizing invasiveness. However, current methods for fabricating miniature fiber probes have limited ability to correct optical aberrations, leading to suboptimal imaging performance. In this study, we introduce a new paradigm of liquid shaping technique for the rapid and scalable fabrication of ultrathin and high-performance OCT microendoscopes suitable for minimally invasive clinical applications. This technique enables the flexible customization of freeform microlenses with sub-nanometer optical surface roughness by regulating the minimum energy state of curable optical liquid on a wettability-modified substrate and precisely controlling the liquid volume and physical boundary on a substrate. Using this technique, we simultaneously fabricated 800-nm OCT microendoscopes with a diameter of approximately 0.6 mm and evaluated their ultrahigh-resolution imaging performance in the esophagus of rats and the aorta and brain of mice.

Introduction

Endoscopic optical coherence tomography (OCT) is a label-free technology that enables the real-time, three-dimensional (3D) *in vivo* imaging of luminal organs with an imaging depth of 1–3 mm in tissues^{1, 2}. This technique allows the near-histologic quality visualization of tissue microstructures and provides histopathological information. Endoscopic OCT overcomes the limitations of traditional biopsy by enabling volumetric sampling across a large area without necessitating tissue removal³. Most current endoscopic OCT systems typically operate at 1,300 nm and have a limited resolution of approximately $10 \mu\text{m}^{3-7}$. To address these limitations, endoscopic OCT operating at 800 nm has been developed. This OCT offers a considerably high

resolution (approximately 2 to 4 μm) and enhances image contrast, albeit at the expense of a shallower imaging depth compared to the 1300-nm OCT⁸⁻¹¹. This technology has been demonstrated to be potentially valuable for evaluating the airways^{12, 13}, gastrointestinal tract^{9, 11, 14}, blood vessels¹⁵, and cervixes¹⁶⁻¹⁸, et al.

When imaging small and convoluted luminal organs, such as narrow arteries and peripheral bronchioles, a miniaturized OCT endoscope that offers ease of access and adequate mechanical flexibility without causing discomfort or damage to tissues is required^{9, 13, 19}. A small probe can extend the limited imaging depth of OCT, facilitating minimally invasive interstitial imaging in solid tissues and organs, such as the brain²⁰. Therefore, an 800-nm OCT endoscope with an ultrahigh resolution (e.g., an axial resolution of $<3 \mu\text{m}$ in the air) and ultrathin form factor (e.g., an outer diameter of $<1 \text{ mm}$) is required for accurately detecting subtle pathological changes in tissues while ensuring minimal invasiveness during imaging.

To date, most miniature OCT probes have been fabricated using all-fiber distal focusing optics, typically composed of a delivery fiber spliced with a graded-index (GRIN) fiber and an angle-polished, side-deflecting fiber mirror²¹ or a fiber ball lens directly melted and angle-polished on the fiber tip⁹. However, a conventional OCT microendoscope composed of GRIN fiber suffers from severe chromatic aberration at 800 nm. Previous achromatic design using a diffractive lens suffers from relatively low transmission efficiency^{22, 23}. Although an achromatic 800-nm OCT microprobe can be developed using an angle-polished distal fiber ball lens, the fiber-melting technique does not provide adequate flexibility to fabricate a ball lens with optimal balance among working distance, resolution, and depth of focus (DOF)^{9, 20}. Additionally, the fabrication of a beam reflector in a fiber ball-lens-based OCT probe involves the use of an angle-polishing procedure that is labor-intensive and presents challenges in achieving optimal optical surface roughness. Recently, two-photon 3D microprinting has been employed to construct freeform side-deflecting micro-optics on a single-mode fiber (SMF), resulting in a 1,300-nm OCT endoscope with a diameter of approximately 0.5 mm¹⁹. However, this microprinting technique is expensive and lacks potential for scalability. Moreover, the surface roughness of 3D-printed optics is approximately 10–200 nm, which is not ideal for OCT imaging^{19, 24}.

In this study, we introduce a novel liquid shaping technique that facilitates the rapid simultaneous fabrication of side-focusing microlenses on fiber tips for use in ultrahigh-resolution OCT microendoscopy. By regulating the minimum energy state of curable optical liquid on a substrate surface with tailored wetting properties, as well as controlling the droplet's volume and its physical boundary on the substrate, the size and shape of the distal microlens of an OCT probe can be flexibly customized using our method. This helps correct chromatic aberration and optimize imaging performance. This technique also eliminates the need for angle-polishing and results in liquid-shaped microlens with sub-nanometer surface roughness. Using this technique, we

successfully fabricated ultrathin, aberration-corrected 800-nm OCT microendoscopes simultaneously. The resulting endoscopes had a diameter of approximately 0.6 mm (including a protective sheath) and provided an ultrahigh resolution of $2.4 \mu\text{m} \times 4.5 \mu\text{m}$ (in axial and transverse directions in air). Furthermore, we demonstrated the imaging performance, mechanical flexibility, and minimal invasiveness of these microendoscopes by imaging the esophagus of rats and the aorta and brain of mice. Our method potentially facilitates the scalable fabrication of cost-effective, high-performance OCT microendoscopes for minimally invasive and ultrahigh-resolution optical biopsies in clinical settings.

Results

Liquid shaping technique




Fig. 1 Schematic illustration of optical microlens fabrication using liquid shaping technique. **a(i)** Production of microlens on a wettability-modified substrate using a piezoelectrically actuated dispenser with thermal control. **a(ii)** Ultrasmall OCT microendoscope based on liquid-shaped microlens for *in vivo* imaging in small and convoluted luminal organs, such as coronary vasculature. **b** and **c** Schematic of optical liquid on a substrate (without physical boundary). The microlenses of the same liquid volume (V_1) on the substrates of different wettability ($W_1 - W_3$) have a different radius, i.e., $r_1 - r_3$, and contact angle, i.e., $\theta_1 - \theta_3$ (**b**). As the liquid volume increases from V_1 to V_3 on the substrate of the same wettability W_4 , the radius of microlens increases from r_4 to r_6 while maintaining a contact angle

of θ_4 (c). **d** Production of spheroid microlens using a substrate with a circular boundary. The radii and contact angles of microlenses made with different liquid volumes are indicated with r_1 , r_2 , and θ_1 , θ_2 , θ_3 , respectively. **e** Fabrication of ellipsoid microlens on a substrate with an elliptical boundary. The semi-major length, i.e., m_1 , m_2 , m_3 , semi-minor length, i.e., n_1 and n_2 , contact angle on x-z plane from θ_{x1} to θ_{x3} , and contact angle on y-z plane from θ_{y1} to θ_{y3} of the microlenses are illustrated when different liquid volumes are used on the substrate. Please note that $n_1 = m_1$, $n_2 = m_2$, $\theta_{x1} = \theta_{y1}$, and $\theta_{x2} = \theta_{y2}$.

Freeform microlenses can be fabricated by manipulating the minimum energy state of curable optical liquid droplets on a wettability-modified substrate^{25, 26} (Fig. 1a(i)) and used as focusing micro-optics in OCT endoscope of ultrasmall size, i.e., OCT microendoscope, for volumetric optical biopsy in complex luminal organs (see Fig. 1a(ii) and Materials and methods). In our work, a piezoelectrically actuated dispenser was used to precisely control the volume of the optical liquid droplet (NOA 81, Norland Product Inc.). Thermal control was employed by heating the dispenser nozzle to approximately 75 °C to reduce the liquid viscosity from 300 to 30 cps, facilitating easy dispensing of the optical liquid^{27, 28} (see Fig. 1a(i) and Materials and methods). This approach enables the production of lenses of various shapes and sizes within tens of minutes by controlling a substrate's wettability (or surface energy, see Fig. 1b and Materials and methods) as well as the liquid volume (Fig. 1c) and physical boundaries (i.e., circular or elliptical, Fig. 1d and e) on a substrate. Furthermore, the liquid-shaped microlens provides a reflective surface with sub-nanometer surface roughness^{29, 30} (Fig. 1a(i)), eliminating the additional angle-polishing process required in conventional methods based on GRIN fiber and fiber ball lens.




Fig. 2 Shape and size control of the microlens using liquid shaping technique. **a** The contact angles and sizes (in radius and height) of microlenses (illustrated in cross-sections) regulated by the wettability of glass substrate (without physical boundary) while using a constant liquid volume of about 50 nL. **b** The microlenses of constant 90° contact angle (in cross-sectional view) with their sizes (in radius and height) controlled by the droplet volumes from 5 nL to 800 nL. **c** Comparison of measured and theoretically calculated radii of microlenses regulated by the liquid volume on the substrates of different wettability (indicated with colors). The green triangle represents the same set of measurements shown in **b**. The insets show three representative microlenses of contact angles and radii of about 50°/440 μm , 75°/334 μm , and 90°/277 μm , respectively. Scale bars are 250 μm . **d** Cross-sectional profiles of spheroid microlenses made on a substrate of a circular boundary (i.e., a radius of 493 μm). **e** Cross-sectional profiles of ellipsoid microlenses fabricated on a substrate of an elliptical boundary (i.e., a semi-major length of 460 μm and a semi-minor length of 360 μm), the shapes and sizes of microlenses are displayed on both x-z (colored solid lines) and y-z planes (colored dashed lines). **f** Study of the shrinkage effect in microlenses (in terms of contact angle and volume) by comparing their cross-sectional profiles before (colored solid lines) and after (colored dotted lines) ultraviolet light-induced polymerization. The shrinkage ratio is calculated using $(V_1 - V_2)/V_1$, where V_1 and V_2 are the lens volume before and after polymerization.

The versatility of this technique was validated through the continuous adjustment of the contact angle of approximately 50 nL of optical liquid on a glass substrate (which has no physical boundary for the liquid) from about 50° to 110° at an angle precision of about 0.5°. This was achieved by

altering the substrate's wettability using a previously reported fluorination-based surface wettability modification method^{31, 32} (see Fig. 2a and Materials and methods). Once the desired contact angle (such as 90°) of the microlens was confirmed, the lens radius ranging from 120 to 720 μm was achieved at a resolution of approximately 1 μm by precisely controlling the droplet volume with a precision of approximately 0.1 nL (see Fig. 2b and Materials and methods), thereby verifying the scale-invariant nature of the liquid shaping technique. The measured radii (and heights) of microlenses made of different liquid volumes on glass substrates with defined wettability and contact angles closely aligned with the theoretically calculated radii³³ (Fig. 2c). This indicates the satisfactory controllability of the liquid shaping technique in creating spherical and aspherical microlenses of predefined shapes and sizes on substrates without physical boundaries.

A physical boundary on a substrate can be utilized to fabricate lenses with other 3D shapes (Fig. 1d and e). For instance, a spheroid lens can be fabricated using a 3D-printed circular cylinder substrate (see Materials and methods and Movie S1). The radius of the liquid lens and contact angle are initially governed by the substrate's wetting property and liquid volume (see the blue, purple, and red solid lines in Fig. 2d) and subsequently determined by the physical boundary size in conjunction with liquid volume (see the orange, green, and black solid lines in Fig. 2d).

The fabrication of an ellipsoid lens was validated using a 3D-printed elliptical cylinder substrate (see Materials and methods and Movie S2). With the increase in liquid volume, the shape and size of the lens are initially controlled by the substrate's wetting property and liquid volume (see the blue, purple, and red solid lines on x-z plane and blue and purple dashed lines on y-z plane in Fig. 2e) and then primarily by the physical boundary constraint (see the orange, green, and black solid lines on x-z plane and the red, orange, green, and black dashed lines on y-z plane in Fig. 2e).

The polymerization of microlenses was performed in approximately 30 minutes using an ultraviolet (UV) lamp operating at a wavelength of around 365 nm and a power density of 65 mW/cm^2 (on a lens sample). A constant shrinkage ratio of 7.69% in volume was identified for the fully polymerized microlenses, which kept almost unchanged contact angle and lens profile as those before polymerization^{34, 35} (see Fig. 2f and Materials and methods). Therefore, the calculated volume of the designed microlens was increased by 8.33% to compensate for the shrinkage effect and thus achieve the desired shape and size of the polymerized microlens (Supplementary Note 1).

Liquid-shaped OCT microendoscope




Fig. 3 Liquid-shaped microendoscope. **a(i)** Fabrication of liquid-shaped OCT fiber probe. Optical liquid of pre-calculated volume on a glass substrate with a modified wetting property forms a microlens of custom contact angle, shape, and size, which is then polymerized and used as distal optics of fiber probe made of single-mode fiber (SMF) and non-core fiber (NCF). Here an incident angle θ is formed in between the OCT light axis with respect to the normal axis of the reflective surface of microlens. **a(ii)** Schematic of OCT microendoscope. The fiber probe is guarded in a hypodermic tube (rigid version) or a torque coil (flexible version) and protected with a glass capillary tube (rigid version) or a plastic sheath (flexible version). **a(iii)** Zoomed-in view of the distal part of the microendoscope, consisting of an NCF of length L and a semi-spherical microlens of radius r and utilizing an incident angle θ . Here the working distance (WD) is measured as the normal distance between the protective sheath surface and the center of the focal plane. **b** Calculated back-reflection in microendoscope versus NCF length L for four representative combinations of lens radius r and incident angle θ . The back-reflection is found to decrease by about 0.55 dB (black line), 0.33 dB (red line), 0.89 dB (blue line), and 0.41 dB (green line), respectively, as NCF length L increases from 350 μm to 650 μm for each combination. **c** Calculated back-reflection versus lens radius r and incident angle θ in the microendoscope using an NCF length of 500 μm . The desired incident angle of 52.5° was indicated by a red dot. **d-h** Chromatic focal shift (**d**), focused spot size (**e**), astigmatism ratio (**f**), effective depth of focus (DOF, **g**), and working distance (**h**) calculated at different NCF lengths L and lens radii r when the incident angle is 52.5°. **i** Optimal design region compiled from the calculated results in **d-h**. The design adopted to fabricate the microendoscope is indicated with a red dot.

The liquid shaping technique facilitates the fabrication of an ultrathin OCT fiber probe, which consists of an SMF, a non-core fiber (NCF), and a custom liquid-shaped microlens (Fig. 3a(i)).

Initially, an NCF piece is spliced to an SMF and then coupled to a polymerized microlens at an incident angle θ by using NOA 81 as an optical adhesive (see Materials and methods). Subsequently, the OCT probe is protected with a hypodermic tube (rigid version) or a torque coil (flexible version). The entire probe is then encased in a protective sheath, such as a glass capillary tube (for the rigid version) or a plastic sheath (for the flexible version), to form an OCT microendoscope (Fig. 3a(ii)). This technique enabled the scalable production and fabrication of five microendoscopes simultaneously (see Materials and methods and Fig. S1).

To demonstrate the feasibility of the liquid shaping technique for imaging performance optimization and aberration correction, we designed an 800-nm OCT microendoscope using a simple semi-spherical microlens (Fig. 3a(iii)). The light incident angle was initially optimized to minimize light back-reflection in the microendoscope by performing stray light analysis in OpticStudio (see Materials and methods). We used the total internal reflection with an incident angle above the critical angle (i.e., approximately 40.2° at 842 nm) on the reflective surface of the microlens. The simulation results indicated that back-reflection in the OCT probe was mainly determined by the incident angle and lens radius, whereas its dependence on the NCF length was relatively low (Fig. 3b). Generally, a larger lens radius and incident angle result in lower back-reflection in the OCT probe. Because a small microlens is preferred to fabricate a miniature probe and a larger incident angle leads to a smaller numerical aperture and larger focused spot size at a fixed working distance, we selected 52.5° in our design to achieve a back-reflection of less than -56 dB for a lens radius from 125 to $150\ \mu\text{m}$ (Fig. 3c).

To further optimize imaging performance, we calculated the chromatic focus shift (within a spectrum ranging from 750 to 950 nm), focused spot size (the average spot size measured in x and y directions on the focal plane), astigmatism ratio (the ratio of spot sizes measured in x and y directions on the focal plane), effective DOF¹⁹, and working distance (WD) for different combinations of NCF length and lens radius using OpticStudio (Fig. 3d-h and Materials and methods). It is noted that the source of astigmatism is derived from the cylindrical wall of the glass capillary or plastic sheath, which has different curvatures parallel and perpendicular to the endoscope axis^{19, 36-38}. Compiling the simulated results indicated an optimal design region (Fig. 3i) that provides optimal imaging performance, such as a chromatic focal shift of less than $6\ \mu\text{m}$, a focused spot size (i.e., transverse resolution) of less than $6\ \mu\text{m}$, an astigmatism ratio between 0.9 and 1.1, an effective DOF of larger than $150\ \mu\text{m}$, and a WD of between 200 and $300\ \mu\text{m}$. We used a design that involves fabricating a liquid-shaped microendoscope using a $500\text{-}\mu\text{m}$ -long NCF, a microlens with a $140\text{-}\mu\text{m}$ radius, and an incident angle of 52.5° . This approach enables achromatic performance to be achieved with a minimal focal shift of approximately $5.4\ \mu\text{m}$, a high transverse resolution of approximately $4.6\ \mu\text{m}$, a low astigmatism ratio of approximately 1.05 on the focal

plane, a low back-reflection of less than -56 dB, and an appropriate DOF and WD of approximately 197.3 and $238 \mu\text{m}$, respectively (Fig. 3i).

Characterization of the OCT microendoscope




Fig. 4 OCT microendoscope characterization. **a** Photograph of a rigid microendoscope and zoomed-in view of its distal optics (inset). **b** Photograph of a flexible microendoscope and zoomed-in view of its distal optics (inset). **c** Measured 3D profile of the liquid-shaped microlens used in microendoscope. **d** 1D surface profiles of microlenses extended by nonlinear least square fitting the measurements along four representative azimuthal angles at 0° , 45° , 90° , and 135° (dashed lines in **c**). **e** Representative surface roughness measured on the top curved surface of the microlenses (dashed circle in **c**). **f** Representative surface roughness measured on the flat reflective surface of microlenses. **g** Representative focused laser beam profile and focused laser spot (inset) measured out of a 0.6-mm protective glass capillary tube of the rigid microendoscope. The x-axis represents the distance relative to the focal plane along the light-emitting direction, and $0 \mu\text{m}$ is the location of focal plane. The beam diameters in x and y directions are measured using an optical beam profiler, and the mean diameters are calculated using weighted equations provided by DataRay, i.e., $w = 0.83114 \times x + 0.16886 \times y$ when $x \geq y$, or $w = 0.16886 \times x + 0.83114 \times y$ when $x < y$; w is the mean diameter, and x and y indicate the beam diameters in x and y directions, respectively. **h** Axial resolution measured along the imaging depth with a representative point spread function (inset). Here $0 \mu\text{m}$ in imaging depth is where the zero optical-delay position between the reference and sample arms is located. **i** Back-reflected spectra obtained by moving

a mirror along the light-emitting direction and measured out of the protective sheath of microendoscope. Here $0\ \mu\text{m}$ indicates the position of focal plane.

Both the rigid and flexible versions of liquid-shaped 800-nm OCT microendoscopes, with diameters of approximately 0.6 mm (including a protective sheath), were fabricated using the aforementioned probe design (see Fig. 4a and b and Materials and methods). Five microendoscopes were simultaneously fabricated in 90 minutes, with comparable results, demonstrating the scalability of the proposed method (see Materials and methods, Supplementary Note 2, and Figs. S1 and S2). To illustrate the fabrication quality and controllability of the liquid shaping technique, the surface profile of the distal semi-spherical microlens was first characterized using a 3D noncontact confocal surface profiler (MarSurf CM Expert, Mahr Inc.). Because the profiler's imageable slope angle was limited to only approximately 45° , we measured only the top $40\text{-}\mu\text{m}$ height of the lens (Fig. 4c). The fitted one-dimensional surface profiles at four azimuthal angles (i.e., 0° , 45° , 90° , and 135°) indicated a microlens of a symmetric semi-spherical shape with a radius of approximately $140.5 \pm 0.5\ \mu\text{m}$ (close to the designed radius of $140\ \mu\text{m}$), suggesting the precise control of lens shape and size (Fig. 4d). Using a white-light interferometry-based surface profiler (MarSurf WI 50, Mahr Inc.), we measured the surface roughness of the microlens on the top curved surface to be $0.84 \pm 0.11\ \text{nm}$ (Fig. 4e). This sub-nanometer roughness is due to the smoothness of the liquid-air interface in the liquid shaping technique^{29,30}. Meanwhile, we noted a surface roughness of $0.53 \pm 0.11\ \text{nm}$ on the flat reflective surface of the microlens, owing to the use of an ultra-flat glass substrate (see Fig. 4f and Materials and methods). A liquid-shaped microlens with sub-nanometer surface roughness enables the effective mitigation of unwanted light scattering in the distal focusing optics of the OCT microendoscope.

To further characterize microendoscopes, we constructed an 800-nm spectral-domain OCT (SD-OCT) system. The configuration of this constructed system was similar to those developed previously^{8,9,20} (see Supplementary Note 3 and Fig. S3). The diameter of the OCT laser beam exiting the protective sheath, such as the glass capillary tube used in the rigid microendoscope, was measured in both x and y directions along the light-emitting direction using an optical beam profiler (BladeCam2-XHR, DataRay Inc., Fig. 4g). The smallest focused spot was measured approximately $240\ \mu\text{m}$ away from the outer sheath surface, exhibiting spot sizes of about 4.5 and $4.3\ \mu\text{m}$ in x and y directions, respectively (with a mean diameter of $4.5\ \mu\text{m}$, inset of Fig. 4g), indicating a low astigmatism ratio of 1.05 on the focal plane. These measurements align well with the simulated focused spot size of $4.6\ \mu\text{m}$ at a working distance of $238\ \mu\text{m}$. Furthermore, the measured effective DOF was approximately $200\ \mu\text{m}$, which was estimated by polynomial-fitting the mean beam diameters within the depth range where the beam diameter was smaller than twice the size of the focused spot¹⁹. Additionally, the achromaticity of the microendoscope was verified based on the observation of less than 5% variation in the axial resolution of approximately $2.43\ \mu\text{m}$ along the 1-mm imaging depth (Fig. 4h) and confirmed by the nearly unchanged back-reflected

spectra measured by moving a mirror along the light-emitting direction⁹ (Fig. 4i). These results demonstrated the ultrahigh resolution of the liquid-shaped OCT microendoscope operating at 800 nm.

Ultrahigh-resolution imaging of small luminal organs




Fig. 5 Imaging rat esophagus using flexible microendoscope. **a** Cut-way view of a reconstructed 3D OCT image of a 36-mm-long rat esophagus. **b** Representative 2D OCT image corresponding to the cross-section boxed with green dashed lines in **a**. **c** 3x close-up view of the region labeled with red dashed box in **b**. **d** Correlated hematoxylin and eosin (H&E) histology. EP: stratified squamous epithelium, LP: lamina propria, MM: muscularis mucosae, SM: submucosa, CM: circular muscle, LM: longitudinal muscle. Scale bars are 250 μm .

Our flexible microendoscope is advantageous in terms of its ultrathin size, mechanical flexibility, and ultrahigh resolution. It can pass through the constricted lumen, such as narrowed sections in the blood vessels, and smoothly scan small luminal organs, such as an infant's esophagus. Our microendoscope enables the detection of fine microstructures and subtle pathologies of diseased tissues *in vivo*. In this study, a rat's esophagus was imaged to evaluate the imaging performance of the microendoscope. The probe initially traveled through the oral cavity of the rat, passed the pharynx, traversed the tight upper esophageal sphincter (a narrow luminal structure that facilitates swallowing and reduces food backflow into the pharynx), and finally reached the small esophagus. 3D volumetric imaging was performed at a speed of 10 frames/second over a 36-mm-long esophagus. A separation of adjacent frames, i.e., pitch number, of 20 μm was used to control the pullback speed of the microendoscope.

The reconstructed 3D volumetric image and the representative OCT cross-section clearly revealed the layered tissue structures of rat's esophagus (Fig. 5a and b). The fine laminar structures of the esophagus, including the stratified squamous epithelium (EP), lamina propria (LP), muscularis mucosae (MM), submucosa (SM), circular muscle (CM), and longitudinal muscle (LM), were clearly observed in a zoomed-in view (Fig. 5c). The microstructures of the esophagus observed on the OCT image aligned with the corresponding hematoxylin and eosin (H&E) histology micrograph (Fig. 5d). Compared with conventional 1,300-nm OCT endoscopes, our liquid-shaped

OCT microendoscope operating at 800 nm enables the acquisition of ultrahigh-resolution images of the fine microstructures in the esophagus. This capability holds the potential for detecting subtle pathologies associated with early-stage diseases *in vivo*³⁹.

In situ imaging of narrow lumens in complex internal organs




Fig. 6 *In situ* imaging of mouse aorta with flexible microendoscope. **a** A flexible microendoscope is deployed from the left femoral artery to the aortic arch in the mouse. **b** Cut-way view of a reconstructed 3D OCT image of a 14.6-mm-long mouse descending thoracic aorta. **c** Representative OCT cross section indicated with a green dashed box in **b**. **d** and **e** 3x close-up view of the region boxed with red dashed lines in **c** and its correlated hematoxylin and eosin (H&E) histology. The white arrow points to the deep tissue. **f** and **g** Example quantification of each fine tissue layer's thickness (distance between red dashed lines) of mouse aorta along an A-line depth (the white arrow in **d**) (**f**). The thickness and its standard deviation of each aorta layer were measured using 40 A-lines from 10 representative OCT cross-sections (**g**). TI: tunica intima, TM: tunica media, TA: tunica adventitia, EL: elastic lamellae, AT: adipose tissue. Scale bars are 100 μm .

Currently, the accurate assessment of high-risk arterial diseases, such as atherosclerosis^{40, 41}, in small vessels remains challenging due to their narrow lumens, highly complex networks, and vast distribution⁴². Thus, a flexible microendoscope that allows for minimally invasive and ultrahigh-resolution imaging in small blood vessels is highly desirable. To evaluate the functionality of our microendoscope in the narrow lumens of blood vessels, we performed the *in situ* imaging of the descending aorta, which has a lumen diameter of approximately 0.6 mm, in a normal mouse model.

In this study, the mouse was first euthanized by administering an overdose of ketamine and xylazine before imaging. The mouse heart and connected vessels were perfused with phosphate-buffered saline using a 27-gauge needle inserted into the apex of left ventricle^{19,43}. This perfusion procedure depleted blood in the aorta, thereby preventing any interference from high-scattering red blood cells during OCT imaging. Subsequently, the microendoscope was inserted through the left femoral artery. It passed through the descending abdominal aorta and thoracic aorta sections, finally reaching the section close to the aortic arch (Fig. 6a).

A 14.6-mm-long section of the descending thoracic aorta was imaged at a speed of 10 frames/second with a frame pitch of 20 μm (Fig. 6b). The imaged aorta was then excised, fixed, embedded, sectioned into 10- μm -thick slices, and stained with H&E. As demonstrated in the OCT cross-section, the zoomed-in view, and the corresponding histology micrograph, we clearly observed the laminar microstructures of the mouse aorta, including the tunica intima (TI), tunica media (TM), tunica adventitia (TA), and adipose tissues (AT) (Fig. 6c-e). Furthermore, multiple elastic lamellae (EL, with high scattering) intermixed with smooth muscle sheets were identified in the TM layer.

The ultrahigh-resolution imaging facilitated by the microendoscope enabled the clear delineation and accurate quantification of the microstructures of the aorta for the *in situ* evaluation of arterial diseases. A preliminary study revealed that in a normal mouse, the thicknesses of the aortic layers were 11.0 ± 1.2 , 69.2 ± 3.4 , and 28.2 ± 2.2 μm for the TI, TM, and TA, respectively (Fig. 6f). In particular, the TI, which is the thinnest and innermost layer of an artery or vein, is composed of a single layer of endothelial cells; its thickening and proliferation are considered an early indication of atherosclerosis^{44, 45}.

Minimally invasive interstitial imaging in deep brain *in vivo*




Fig. 7 *In vivo* imaging of mouse deep brain with rigid microendoscope. **a** The preparation of mouse brain for imaging with the scalp incised and two burr holes (about 1 mm in diameter) on skull. **b** Schematic of microendoscope imaging in mouse deep brain. **c** Reconstructed volumetric mouse brain image of 5-mm depth. **d** *En face* projection view of the unfolded cylindrical OCT volume shown in **c**. **e** Hematoxylin and eosin (H&E) histology micrograph of brain sample correlated to **c** and **d**. **f-i** Representative cross-sectional images of different mouse brain structures, including cerebral cortex (**f**), corpus callosum (**g**), caudate putamen (**h**), and ventral striatum (**i**), corresponding to dashed lines in **c**. Scale bars are 500 μm .

The ultrathin rigid microendoscope can extend the limited imaging depth of OCT and perform high-resolution, volumetric interstitial imaging of solid organs with minimal invasiveness. This facilitates the access and evaluation of deep-seated diseases, such as deep brain tumors, ischemic stroke, and epilepsy⁴⁶⁻⁴⁹, while minimizing the risks of hemorrhage and other trauma caused by probe insertion²⁰.

To evaluate the functionality of the rigid microendoscope, we performed *in vivo* deep-brain imaging in a mouse brain. After making an incision on the scalp, we drilled two small burr holes on each of the two contralateral sides of the skull (Fig. 7a). The microendoscope was then inserted through the burr holes into the deep brain, following an insertion trajectory perpendicular to the brain surface (Fig. 7b). A 5-mm-long (in *z* direction) cylindrical deep brain volume was imaged in 50 seconds at a speed of 10 frames/second. Because of the ultrahigh resolution of the microendoscope, we clearly observed mouse brain structures, including the cerebral cortex, corpus callosum, caudate putamen, and ventral striatum, on the 3D volumetric image (Fig. 7c). The *en face* projection view, generated by summing the unfolded cylindrical brain volume along the

imaging depth, revealed the distinct laminated brain structures of the mouse, which were also verified in the histology micrograph, indicating a good correlation between the OCT representation and histomorphology (Figs. 7d-e and S4). Representative OCT cross-sections acquired at different depths illustrate the detailed morphological features of each mouse brain structure (Fig. 7f-i). In particular, the filament bundle structures of striatopallidal fibers in the caudate putamen were clearly observed on the OCT images²⁰ (Fig. 4c, d, and h).

Discussion

In comparison to the 1300-nm OCT, the 800-nm OCT provides improved imaging resolution and contrast, albeit with a reduced imaging depth. As conventional GRIN-fiber-based miniature OCT endoscopes experience severe chromatic aberration at 800 nm, we previously developed an achromatic microprobe with a diameter of 1 mm by utilizing a monolithic fiber ball lens with the fiber melting technique⁹. However, to further minimize the probe size to below 1 mm, the current fabrication technique utilizing the fiber melting process is suboptimal, yielding a small fiber ball lens with a low resolution and short DOF⁹. In addition, the miniature probe has a limited numerical aperture, which causes difficulty in achieving an optimal balance between high resolution and DOF.

An OCT microendoscope with a high resolution and large DOF is necessary for the accurate imaging of subtle pathological changes in tissues. For this purpose, an appropriately designed microlens is required in the microendoscope. However, it is difficult to precisely tailor a microlens by using the fiber melting technique. To overcome this limitation, we propose the liquid shaping technique. This method allows for the precise customization of the shape and size of a liquid droplet on a substrate, thereby producing a freeform microlens with an ultrasmall form factor, corrected aberrations, and desired imaging performance. These features can be achieved by conveniently modifying the wettability of a substrate, controlling the liquid volume, and utilizing physical constraints on the substrate. In contrast to the 3D printing method, our technique does not require expensive high-end machinery and can be easily scaled up. Additionally, the liquid shaping technique yields a custom microlens with a sub-nanometer surface roughness (Fig. 4), which is considerably better than the roughness (approximately 10-200 nm) achieved using the two-photon 3D microprinting^{19, 24} (see Supplementary Table 1 for comparison).

The liquid shaping technique offers a novel approach to simultaneously fabricate high-performance OCT microendoscopes. The fabrication process of the microendoscope involves only standard and common optical fiber handling procedures, such as splicing and cleaving, and the simple gluing of a custom microlens to the fiber probe tip (see Materials and methods and Fig. S1). These processes can be streamlined for mass production. This approach eliminates the need for a time-consuming angle-polishing process that is required in conventional GRIN-fiber and

fiber ball-lens-based methods. The use of our approach substantially reduces both fabrication time and cost and increases the yield rate due to the minimized reliance on human expertise (see Supplementary Table 1 for comparison). Furthermore, our longitudinal study revealed that the fabricated microendoscopes maintain their imaging performance over a long period (Fig. S5). Thus, the liquid-shaped microendoscope can serve as a low-cost and disposable OCT catheter for translational use.

In principle, the form factor of the microendoscope fabricated using the liquid shaping technique can be further minimized using a thinner hypodermic tube or torque coil and a smaller protective sheath over the entire probe. The imaging performance of a downsized microendoscope can be optimized by customizing a smaller microlens with appropriate back-reflection, achromaticity, astigmatism ratio, resolution, and DOF (Fig. 3). Likewise, the form factor of the microendoscope can be increased to image larger lumens (e.g., those with a diameter greater than 2 mm) by fabricating an appropriately designed microlens. In contrast, it would be suboptimal to fabricate a fiber ball-lens-based microprobe to image lumens larger than 2 mm because a microprobe with a longer working distance tends to lead to a degraded transverse resolution and, more importantly, an increased longitudinal focal shift⁹.

Given that the conventional 1,300-nm endoscopic OCT has successfully gained approval from the Food and Drug Administration for imaging the esophagus and coronary artery, the ultracompact and flexible liquid-shaped microendoscope operating at 800 nm presents promising opportunities for the *in vivo* ultrahigh-resolution imaging of tissue pathological changes in small and/or complex luminal organs, such as small airways and small blood vessels. The ability to obtain ultrahigh-resolution 3D microscopic images of a lumen enhances the likelihood of detecting premalignant or early-stage diseases. In addition, the decreased size and enhanced flexibility of the microendoscope enable sharp bending along the delivery path and substantially reduce the risk when used to image a small and delicate luminal organ. This ultrathin microendoscope can be potentially integrated within a biopsy needle, enabling the accurate assessment of the fine structures of target tissues. This would help guide biopsies, resulting in an improved diagnostic yield.

The liquid-shaped microendoscope is currently in its early stage of development, and substantial technical improvements can be achieved in the future. First, the current liquid shaping technique mainly relies on passive methods, such as substrate wettability, liquid volume, and physical boundary, to control the shape and size of the lens. Active methods that manipulate the thermal and/or electromagnetic field will be explored to gain new control dimensions for shaping microlens with complex freeform surfaces and advanced functionality^{50, 51}. Second, the current study demonstrated only the feasibility of fabricating five liquid-shaped microendoscopes simultaneously. We plan to further streamline and automate the fabrication process to increase the

yield rate and reduce cost. In addition, we validated the imaging performance of the liquid-shaped microendoscope in small animals only. A systematic study of large animals is imperative to thoroughly validate the functionality and safety of our microendoscope before initiating clinical trials with patients. For future clinical use, the imaging speed of the microendoscope should be improved to minimize motion artifacts and achieve a short intervention time in the operating room. It is important to highlight that the proposed liquid shaping technique can be easily implemented in other types of fiberscopes that require a microlens on the fiber tip. This includes confocal, two-photon, and coherent Raman endoscopes, among others⁵²⁻⁵⁴.

Materials and methods

Liquid droplet generation and characterization

A piezoelectrically actuated dispenser (SA306, Sans Inc.) was utilized for droplet/microlens generation. This technique was reported in previous works and widely used for microarray printing⁵⁵, tissue engineering²⁸, and fabrication of functional materials²⁷. To account for the system dispensing accuracy, the calibration of the dispensed liquid volume on the substrate was first performed (Fig. S6). A series of liquid volumes claimed by the dispenser was dispensed on the glass substrate to form the liquid microlens; the contact angle, dimensions, and volume of microlens were then measured using a liquid drop analysis system (OCA 25, Dataphysics Instruments GmbH) at room temperature of 25 °C and analyzed with SCA 20 software (DataPhysics Instruments GmbH). After calibration, the liquid microlens volume can be controlled at a precision of about 0.1 nL. The accuracy of contact angle measurement is $\pm 0.1^\circ$, the accuracy of length measurement is $\pm 0.5 \mu\text{m}$, and the accuracy of volume measurement is $\pm 0.05 \text{ nL}$.

As for the polymerized microlens, the contact angle, dimensions, and volume were characterized using the same method as liquid microlens, while the lens profile was measured using a confocal surface profiler (MarSurf CM Expert, Mahr Inc.) and its surface roughness was characterized with a white-light interferometry-based surface profiler (MarSurf WI 50, Mahr Inc.).

Surface wettability modification of glass substrate

Glass substrates (S2006A1, Ossila) used in our study provide a super-polished surface of about 1-nm roughness. To modify the surface wettability, the glass substrate is first treated in oxygen plasma bathing for 10 minutes. The substrate is then placed in a petri dish and soaked in the mixture of 1H, 1H, 2H, 2H-Perfluorooctyltriethoxysilane (POTS) (volume: 10-20 μL) and methylbenzene (volume: 10 mL). The petri dish is sealed and placed in a fume hood at room temperature for 4 to 16 hours to achieve desired surface wettability and contact angle on the substrate (see Supplementary Table 2). After that, the glass substrate is retrieved and thoroughly rinsed with absolute ethanol before use.

The fabrication and surface wettability modification of cylinder substrates

A stereolithography 3D printer (S300, nanoArch) was employed to fabricate cylinder substrates of circular or elliptical boundary with an about 400- μm height. Polyethylene Glycol Diacrylate (PEGDA) was used for 3D printing.

To modify the surface wettability, the cylinder substrate is first processed with oxygen plasma bathing for 10 minutes. Then, it is placed in a petri dish with the cylinder top surface immersed in 10 μL of POTS. The petri dish is sealed with Kapton tape and placed in a thermotank with a baking temperature of 100°C for 4 to 8 hours to achieve desired surface wettability and contact angle on the substrate (see Supplementary Table 3). After that, the cylinder substrate is retrieved and thoroughly rinsed with absolute ethanol before use.

The fabrication of liquid-shaped microlens

The critical procedures to fabricate microlens using liquid shaping technique are illustrated in Fig. S1a and b. Specifically, NOA 81 was selected over other optical liquids in this study (see Supplementary Note 4 for rationale). The NOA 81 liquid was first degassed in a vacuum chamber (98 kPa vacuum level for 10 minutes) to remove any potential micro-bubbles. Then, the liquid of calculated volume was dispensed using the piezoelectrically actuated dispenser (SA306, Sans Inc.) on wettability-modified glass substrate or circular/elliptical cylinder substrates. The desired shape of size of microlens is obtained by precisely controlling liquid volume, wettability, and physical boundary of substrate.

After that, the liquid microlens on substrate was degassed again in a vacuum chamber (98 kPa vacuum level for 10 minutes). Finally, the liquid microlens was polymerized by using a UV lamp illumination for 30 minutes to ensure complete polymerization, which usually only needs 20 minutes. The UV lamp has a center wavelength of about 365 nm and provides power density of 65 mW/cm^2 on the lens sample. It should be noted that a polymerization-induced shrinkage effect of optical liquid (about 7.69%, see Fig. 2f and Supplementary Note 1) was considered and pre-compensated in volume calculation during the microlens fabrication. For the microendoscopes demonstrated in current work, the lens liquid with a volume of about 6.2 nL was dispensed on the substrate to fabricate the semi-spherical microlens of about 140- μm radius and contact angle of about 89.8°.

Zemax simulations

OpticStudio (v17, Zemax LLC) was used to design and optimize the liquid-shaped microendoscope. The refractive index profiles and material dispersions of silica (of NCF) and NOA 81 are considered in simulations (Fig. S7). The back-reflection under different incident angles, lens radii, and NCF lengths was first calculated using stray light analysis in the non-

sequential mode. After that, the ray-tracing simulation of the microendoscope was performed in the mixed sequential/non-sequential mode. By considering a single-mode fiber with numerical aperture of 0.13 and a spectrum ranging from 750 nm to 950 nm, the chromatic focal shift, focused spot size, astigmatism ratio, effective DOF, and working distance were calculated under different lens radii and NCF lengths.

The simultaneous fabrication of microendoscopes

The key procedures to fabricate microendoscopes are illustrated in Fig. S1c and d.

Essentially, a single-mode fiber (780HP, Thorlabs Inc.) was first spliced with a non-core fiber (FG125LA, Thorlabs Inc.), which was then cleaved to a length of $500 \pm 5 \mu\text{m}$, to form a fiber probe. Both fiber splicing and cleaving procedures were performed using the same glass processor (GPX3800, Thorlabs Inc.) and five fiber probes were prepared in this step.

Second, the polymerized microlenses were coupled to fiber probes at an incident angle of 52.5° . Five microendoscopes were fabricated simultaneously with the homemade four-dimensional assembly setup (Fig. S1e). To ensure precise alignment between the fiber probe and microlens, we developed a high-precision four-dimensional assembly stage equipped with two inspection microscopes that offer top and side views. This stage initially establishes a predetermined incident angle between the fiber probe and microlens. Subsequently, the fine adjustment of the fiber-lens incident angle and distance is achieved with the assistance of the two microscopes. Following this, approximately 0.5 nL of NOA 81 is applied between the fiber tip and microlens to facilitate optical bonding, which is then cured using UV light. The fiber probe tip kept a safe distance of about $5 \mu\text{m}$ to microlens surface during the bonding procedure. 30-minute UV light illumination (wavelength: 365 nm, power density on sample: $65 \text{ mW}/\text{cm}^2$) was used to achieve a complete polymerization of the bonding adhesive.

Third, five fiber probes with a bonded microlens were removed gently from the glass substrate with the aid of a razor blade. The resulting fiber probes usually afford a one-way transmission efficiency of at least 94%.

Finally, the 215-mm-long fiber probes were encased in 26-gauge hypodermic tubes of an open window at the end to make rigid microendoscopes. Silica glass capillary (Molex LLC) of about $617 \mu\text{m}$ in outer diameter and about $40 \mu\text{m}$ in wall thickness was employed to protect the microendoscope during imaging.

Flexible microendoscopes of 215 mm in length were fabricated using the similar procedures. After removing the fiber probes from the glass substrate, torque coils of $450 \mu\text{m}$ in outer diameter were then used to encase the imaging probes. The flexible microendoscope was protected with a fluorinated ethylene propylene (FEP) plastic sheath (Zeus Inc.) of a $626\text{-}\mu\text{m}$ outer diameter and a

45- μm wall thickness during imaging. The total fabrication time of the microendoscope was about 90 minutes.

Animal studies and histological correlation

The protocol of OCT endoscopic imaging in rat and mouse was approved by the Laboratory Animal Services Centre at The Chinese University of Hong Kong.

For rat esophagus ($n = 4$, Sprague Dawley rats) and mouse aorta ($n = 4$, nude mice) imaging, animals were euthanized by overdosing of ketamine (100 mg/kg) and xylazine (16 mg/kg) before OCT imaging. The flexible microendoscope was first deployed near to rat GEJ section and mouse aortic arch, respectively. Then, OCT pullback imaging was performed at a speed of 10 frames/second. The microendoscope probe was retreated after imaging, while the plastic sheath was left in the lumens for registration of the imaged tissues. The imaged esophagus and aorta were harvested and fixed in formalin together with plastic sheath overnight before being submitted for histological processing. Standard H&E slides were obtained and correlated with OCT imaging results.

For deep brain imaging ($n = 4$, nude mice), the mouse anesthetization was first introduced by intraperitoneal injection of ketamine (100 mg/kg) and xylazine (16 mg/kg) and further maintained by inhaling 2% isoflurane with medical oxygen. Burr holes (about 1 mm in diameter) were made on two contralateral sides of mouse skull to allow the deployment of rigid microendoscope (Fig. 7a). The burr hole location was selected to avoid major blood vessels. Before insertion, a thin needle (with a diameter of about 300 μm) was utilized to introduce a small hole through the pia mater. The microendoscope was then inserted through the burr hole into the mouse brain with a slow speed of about 10 $\mu\text{m/s}$. After imaging, the mouse was sacrificed and the brain was immediately harvested. A glass capillary tube was inserted back into the brain tissue along the same imaging trajectory, helping register the imaged tissue. The brain tissue together with glass capillary was then placed in formalin for 48 hours. After fixation, the brain tissue was divided into two sections (Fig. S8) for further histological processing. Standard H&E slides were obtained and correlated with *en face* OCT images.

Data availability

All data needed to evaluate the conclusions in the paper are present in the paper and/or the Supplementary Information. Additional data related to this paper may be requested from the corresponding author.

Reference

1. Fujimoto, J. G. Optical coherence tomography for ultrahigh resolution in vivo imaging. *Nat. Biotechnol.* **21**, 1361-1367 (2003).

2. Israelsen, N. M. et al. Real-time high-resolution mid-infrared optical coherence tomography. *Light Sci. Appl.* **8**, 11 (2019).
3. Adler, D. C. et al. Three-dimensional endomicroscopy using optical coherence tomography. *Nat. Photonics* **1**, 709-716 (2007).
4. Syed, S., Larin, K., Dickinson, M. & Larina, I. Optical coherence tomography for high-resolution imaging of mouse development in utero. *J. Biomed. Opt.* **16**, 046004 (2011).
5. Lin, N. C., Fallah, E., Strimbu, C. E., Hendon, C. P. & Olson, E. S. Scanning optical coherence tomography probe for in vivo imaging and displacement measurements in the cochlea. *Biomed. Opt. Express* **10**, 1032-1043 (2019).
6. Yu, X., Singh-Moon, R. P. & Hendon, C. P. Real-time assessment of catheter contact and orientation using an integrated optical coherence tomography cardiac ablation catheter. *Appl. Opt.* **58**, 3823-3829 (2019).
7. Rao D. S, S. et al. Shot-noise limited, supercontinuum-based optical coherence tomography. *Light Sci. Appl.* **10**, 133 (2021).
8. Yuan, W. et al. Optimal operational conditions for supercontinuum-based ultrahigh-resolution endoscopic OCT imaging. *Opt. Lett.* **41**, 250-253 (2016).
9. Yuan, W., Brown, R., Mitzner, W., Yarmus, L. & Li, X. Super-achromatic monolithic microprobe for ultrahigh-resolution endoscopic optical coherence tomography at 800 nm. *Nat. Commun.* **8**, 1531 (2017).
10. Zhang, W. et al. High-resolution, in vivo multimodal photoacoustic microscopy, optical coherence tomography, and fluorescence microscopy imaging of rabbit retinal neovascularization. *Light Sci. Appl.* **7**, 103 (2018).
11. Yuan, W. et al. In vivo assessment of inflammatory bowel disease in rats with ultrahigh-resolution colonoscopic OCT. *Biomed. Opt. Express* **13**, 2091-2102 (2022).
12. Thiboutot, J. et al. Visualization and validation of the microstructures in the airway wall in vivo using diffractive optical coherence tomography. *Acad. Radiol.*, (2022).
13. Yuan, W. et al. Direct visualization and quantitative imaging of small airway anatomy in vivo using deep learning assisted diffractive OCT. *IEEE Trans. Biomed. Eng.* **PP**, 1-9 (2022).
14. Luo, H. et al. Human colorectal cancer tissue assessment using optical coherence tomography catheter and deep learning. *J. Biophotonics* **15**, e202100349 (2022).
15. Kim, J. et al. Flexible endoscopic micro-optical coherence tomography for three-dimensional imaging of the arterial microstructure. *Sci. Rep.* **10**, 9248 (2020).
16. Zeng, X. et al. Ultrahigh-resolution optical coherence microscopy accurately classifies precancerous and cancerous human cervix free of labeling. *Theranostics* **8**, 3099-3110 (2018).
17. Ma, Y. et al. Computer-aided diagnosis of label-free 3-D optical coherence microscopy images of human cervical tissue. *IEEE Trans. Biomed. Eng.* **66**, 2447-2456 (2019).
18. Ren, C. et al. Multi-center clinical study using optical coherence tomography for evaluation of cervical lesions in-vivo. *Sci. Rep.* **11**, 7507 (2021).
19. Li, J. et al. Ultrathin monolithic 3D printed optical coherence tomography endoscopy for preclinical and clinical use. *Light Sci. Appl.* **9**, 124 (2020).
20. Yuan, W. et al. Theranostic OCT microneedle for fast ultrahigh-resolution deep-brain imaging and efficient laser ablation in vivo. *Sci. Adv.* **6**, eaaz9664 (2020).
21. Keenan, M. et al. Design and characterization of a combined OCT and wide field imaging falloposcope for ovarian cancer detection. *Biomed. Opt. Express* **8**, 124-136 (2017).

22. Mavadia-Shukla, J. et al. High-speed, ultrahigh-resolution distal scanning OCT endoscopy at 800 nm for in vivo imaging of colon tumorigenesis on murine models. *Biomed Opt Express* **9**, 3731-3739 (2018).
23. Li, K. et al. Super-achromatic optical coherence tomography capsule for ultrahigh-resolution imaging of esophagus. *J Biophotonics* **12**, e201800205 (2019).
24. Li, J. et al. 3D-printed micro lens-in-lens for in vivo multimodal microendoscopy. *Small* **18**, e2107032 (2022).
25. Cadarso, V. J. et al. Microlenses with defined contour shapes. *Opt. Express* **19**, 18665-18670 (2011).
26. Roy, A. C., Yadav, M., Arul, E. P., Khanna, A. & Ghatak, A. Generation of aspherical optical lenses via arrested spreading and pinching of a cross-linkable liquid. *Langmuir* **32**, 5356-5364 (2016).
27. Kim, J. Y. et al. Inkjet-printed multicolor arrays of highly luminescent nanocrystal-based nanocomposites. *Small* **5**, 1051-1057 (2009).
28. Pataky, K. et al. Microdrop printing of hydrogel bioinks into 3D tissue-like geometries. *Adv. Mater.* **24**, 391-396 (2012).
29. Elgarisi, M., Frumkin, V., Luria, O. & Bercovici, M. Fabrication of freeform optical components by fluidic shaping. *Optica* **8**, 1501-1506 (2021).
30. Eshel, R. et al. Programmable thermocapillary shaping of thin liquid films. *Flow* **2**, E27 (2022).
31. Ikawa, M. et al. Simple push coating of polymer thin-film transistors. *Nat. Commun.* **3**, 1176 (2012).
32. Ganesh, V. A. et al. Superhydrophobic fluorinated POSS–PVDF–HFP nanocomposite coating on glass by electrospinning. *J. Mater. Chem.* **22**, 18479-18485 (2012).
33. Clarke, A., Blake, T. D., Carruthers, K. & Woodward, A. Spreading and imbibition of liquid droplets on porous surfaces. *Langmuir* **18**, 2980-2984 (2002).
34. Maruno, T. & Murata, N. Properties of a UV-curable, durable precision adhesive. *J. Adhes. Sci. Technol.* **9**, 1343-1355 (1995).
35. Takahashi, A., Sekiguchi, Y. & Sato, C. Volume change and viscoelastic properties of UV-curable adhesives for precise positioning during curing process and their formulation. *The Journal of Adhesion* **98**, 2029-2044 (2022).
36. Xi, J. et al. High-resolution OCT balloon imaging catheter with astigmatism correction. *Opt Lett* **34**, 1943-1945 (2009).
37. Pahlevaninezhad, H. et al. Nano-optic endoscope for high-resolution optical coherence tomography in vivo. *Nat Photonics* **12**, 540-547 (2018).
38. Woo Lee, M., Hoon Kim, Y., Xing, J. & Yoo, H. Astigmatism-corrected endoscopic imaging probe for optical coherence tomography using soft lithography. *Opt Lett* **45**, 4867-4870 (2020).
39. Chen, Y. et al. Ultrahigh resolution optical coherence tomography of Barrett's esophagus: preliminary descriptive clinical study correlating images with histology. *Endoscopy* **39**, 599-605 (2007).
40. Yoo, H. et al. Intra-arterial catheter for simultaneous microstructural and molecular imaging in vivo. *Nat. Med.* **17**, 1680-1684 (2011).
41. Dillon, T. M., Liang, P., Schermerhorn, M. L. & Roche, E. T. A computational program for automated surgical planning of fenestrated endovascular repair. *Communications Engineering* **2**, 37 (2023).
42. Araki, M. et al. Optical coherence tomography in coronary atherosclerosis assessment and intervention. *Nat. Rev. Cardiol.* **19**, 684-703 (2022).
43. Tahara, S. et al. Intravascular optical coherence tomography detection of atherosclerosis and inflammation in murine aorta. *Arteriosclerosis Thrombosis and Vascular Biology* **32**, 1150-1157 (2012).

44. Nakashima, Y., Wight, T. N. & Sueishi, K. Early atherosclerosis in humans: role of diffuse intimal thickening and extracellular matrix proteoglycans. *Cardiovascular Research* **79**, 14-23 (2008).
45. Boua, P. R. et al. Genetic associations with carotid intima-media thickness link to atherosclerosis with sex-specific effects in sub-Saharan Africans. *Nat. Commun.* **13**, 855 (2022).
46. Böhringer, H. J. et al. Time-domain and spectral-domain optical coherence tomography in the analysis of brain tumor tissue. *Lasers Surg. Med.* **38**, 588-597 (2006).
47. Tsytsarev, V., Rao, B., Maslov, K. I., Li, L. & Wang, L. V. Photoacoustic and optical coherence tomography of epilepsy with high temporal and spatial resolution and dual optical contrasts. *J. Neurosci. Methods* **216**, 142-145 (2013).
48. Liu, Q. et al. Monitoring acute stroke in mouse model using laser speckle imaging-guided visible-light optical coherence tomography. *IEEE Trans. Biomed. Eng.* **65**, 2136-2142 (2018).
49. Beckmann, L. et al. Longitudinal deep-brain imaging in mouse using visible-light optical coherence tomography through chronic microprism cranial window. *Biomed. Opt. Express* **10**, 5235-5250 (2019).
50. Lee, S. Y., Chen, W. C., Tung, H. W. & Fang, W. Microlens with tunable astigmatism. *IEEE Photonics Technol. Lett.* **19**, 1383-1385 (2007).
51. Ni, X. et al. Soft shape-programmable surfaces by fast electromagnetic actuation of liquid metal networks. *Nat. Commun.* **13**, 5576 (2022).
52. Kim, J. B., Jeon, J., Hwang, K., Kim, D. Y. & Jeong, K. H. Objective-lens-free confocal endomicroscope using Lissajous scanning lensed-fiber. *J. Opt. Microsyst.* **1**, (2021).
53. Liang, W. X. et al. Throughput-Speed Product Augmentation for Scanning Fiber-Optic Two-Photon Endomicroscopy. *IEEE Trans. Med. Imaging* **39**, 3779-3787 (2020).
54. Lombardini, A. et al. High-resolution multimodal flexible coherent Raman endoscope. *Light-Sci Appl* **7**, (2018).
55. Zhu, X., Zhu, L., Chen, H., Yang, M. & Zhang, W. Fabrication of multi-scale micro-lens arrays on hydrophobic surfaces using a drop-on-demand droplet generator. *Opt. Laser Technol.* **66**, 156-165 (2015).

Acknowledgments

The authors are grateful to Mr. Paul Zhou for his help with the microlens characterizations. This work was supported by the Research Grants Council (RGC) of Hong Kong SAR (ECS24211020, GRF14203821, GRF14216222), the Innovation and Technology Fund (ITF) of Hong Kong SAR (ITS/240/21), and the Science, Technology and Innovation Commission (STIC) of Shenzhen Municipality (SGDX20220530111005039).

Author contributions

W.Y.: Conceived the idea; C.X.: Ran the simulations; designed, fabricated, and characterized the microendoscopes; W.Y., C.X.: Designed the animal experiments; C.X., X.G.: Conducted the surface processing of substrates; C.X., S.A.: Performed the animal experiments and histology; C.X., N.X.: Conducted the 3D printing of cylinder substrates; W.Y.: Supervised the theoretical and experimental work; T. N., L.Z., H.H., S.N.: Contributed to the paper revision; All the authors contributed to the paper writing.

Additional information**Supplementary Information**

Supplementary Information is with this paper.

Competing interests

The authors declare no competing financial interests.

Supplementary Information for Liquid-shaped microlens for scalable production of ultrahigh-resolution OCT microendoscope

Chao Xu¹, Xin Guan², Syeda Aimen Abbasi¹, Neng Xia³, To Ngai², Li Zhang³, Ho-Pui Ho¹, Sze Hang Calvin Ng⁴, and Wu Yuan^{1,*}

¹Department of Biomedical Engineering, The Chinese University of Hong Kong, Hong Kong SAR, China

²Department of Chemistry, The Chinese University of Hong Kong, Hong Kong SAR, China

³Department of Mechanical and Automation Engineering, The Chinese University of Hong Kong, Hong Kong SAR, China

⁴Department of Surgery, The Chinese University of Hong Kong, Hong Kong SAR, China

*wyuan@cuhk.edu.hk

The PDF file includes:

Supplementary Notes 1-4

Figs. S1 to S8

Supplementary Tables 1-3

Other Supplementary Materials available for this manuscript include the following:

Movies S1-2

Contents

Supplementary Information for Liquid-shaped microlens for scalable production of ultrahigh-resolution OCT microendoscope.....	1
Supplementary Note 1: Shrinkage of microlenses before and after polymerization.....	3
Supplementary Note 2: Characterization and imaging results of simultaneously fabricated liquid-shaped microendoscopes.	4
Supplementary Note 3: Ultrahigh-resolution endoscopic SD-OCT system.....	5
Supplementary Note 4: Choice of optical liquids for microlens fabrication.....	6
Figure S1. Procedures for fabricating liquid-shaped microlens and microendoscope.....	7
Figure S2. Imaging results of simultaneously fabricated microendoscopes.	8
Figure S3. Schematic of the endoscopic spectral-domain OCT (SD-OCT) system working near 800 nm.....	9
Figure S4. Illustration of OCT scanning in mouse deep brain and the <i>en face</i> projection procedures.	10
Figure S5. Long-term stability on imaging performance of fabricated microendoscopes.....	11
Figure S6. Calibration of dispensed liquid volume.....	12
Figure S7. Refractive index profile of NOA 81 under room temperature.	13
Figure S8. Schematic of mouse brain handling for histology.	14
Supplementary Table 1: Comparison of three fabrication methods for microendoscopes.	15
Supplementary Table 2: Six groups of glass substrates with different parameters of surface wettability modification.	16
Supplementary Table 3: Two groups of cylinder substrates with different parameters of surface wettability modification.	17
Movie S1: Microlens on a circular cylinder substrate.....	18
Movie S2: Microlens on an elliptical cylinder substrate.....	18

Supplementary Note 1: Shrinkage of microlenses before and after polymerization.

There is a scale-invariable shrinkage of the liquid-shaped micro-lens after the complete polymerization, which keeps the contact angle and shape unchanged. To study the shrinkage ratio of NOA 81-based micro-lens, a 365-nm ultraviolet lamp was used to provide a power density of 65 mW/cm² (on lens sample). The volumes of the micro-lens before (V_1) and after (V_2) polymerization were measured using a liquid drop analysis system (OCA 25, Dataphysics Instruments GmbH) at room temperature of 25°C. As shown in Table 1 below, the average shrinkage ratio in volume (defined as $\frac{V_1-V_2}{V_1}$) is found to be about 7.69% with a standard deviation of 0.39% for NOA 81-based micro-lens.

Table 1: Volume variation of microlenses before and after polymerization.

Volume before exposure (V_1 , nL)	Volume after exposure (V_2 , nL)	Shrinkage ratio ($\frac{V_1-V_2}{V_1}$)	Averaged shrinkage ratio
1841.4	1712.0	7.03%	7.69%±0.39%
1521.7	1393.5	8.43%	
1263.7	1176.0	6.93%	
765.6	712.2	6.93%	
658.5	607.6	7.73%	
470.7	430.1	8.62%	
266.5	244.9	8.12%	
213.3	197.3	7.51%	
149.7	138.1	7.74%	
51.1	46.8	8.41%	
28.6	26.4	7.34%	
10.8	10.0	7.41%	

Therefore, an increase of 8.33% of the calculated volume of the designed microlens is used to compensate for the shrinkage effect to help achieve the desired shape and size of the polymerized microlens. As for the microlens used in microendoscopes, the dispensed liquid volume was about 6.2 nL to achieve a semi-spheroid lens of a radius of 140 μm and lens volume of about 5.7 nL.

Supplementary Note 2: Characterization and imaging results of simultaneously fabricated liquid-shaped microendoscopes.

Five microendoscopes were simultaneously fabricated, including two rigid ones (# 1 and # 2) and three flexible ones (# 3, # 4, and # 5). They were characterized of similar results (see Table 2 below) and tested for OCT imaging (Fig. S2) with comparable image qualities, showing the scalability of the liquid shaping technique for microendoscope fabrication.

Table 2: Characterization of the simultaneously fabricated liquid-shaped microendoscopes.

Characterized parameter	# 1*	# 2	# 3**	# 4	# 5
x focused spot size	4.5 μm	4.6 μm	4.5 μm	4.7 μm	4.5 μm
y focused spot size	4.3 μm	4.4 μm	4.3 μm	4.5 μm	4.2 μm
Mean focused spot size [a]	4.5 μm	4.6 μm	4.5 μm	4.7 μm	4.4 μm
Astigmatism ratio	1.05	1.05	1.05	1.04	1.07
Effective DOF	200 μm	202 μm	198 μm	204 μm	196 μm
Working distance	240 μm	235 μm	241 μm	237 μm	243 μm
Mean axial resolution [b]	2.43 μm	2.52 μm	2.43 μm	2.45 μm	2.51 μm
Axial resolution variation [c]	4.6%	4.7%	4.6%	5.0%	4.8%
Mean spectral variation [d]	0.0136	0.0141	0.0138	0.1315	0.0145

* The reported rigid microendoscope.

** The reported flexible microendoscope.

[a] Mean focused spot size = $0.83114 \times x + 0.16886 \times y$ when $x \geq y$, mean focused spot size = $0.16886 \times x + 0.83114 \times y$ when $x < y$, and x and y indicate the beam diameters in x and y directions, respectively (provided by DataRay Inc.).

[b - c] The axial resolution is measured every 100 μm along the imaging depth of 100 μm to 1000 μm , and mean axial resolution and axial resolution variation can be calculated.

[d] The mean spectral variation (MSV) is calculated by $\text{MSV} = \text{MEAN}(\text{STD}(S_1, S_2, \dots, S_i))$, where S_i is a 2048×1 vector that contains the normalized spectral data, MEAN is the averaging operation, and STD indicates the standard deviation.

Supplementary Note 3: Ultrahigh-resolution endoscopic SD-OCT system.

To characterize the performance of the microendoscope, a homemade endoscopic spectral-domain OCT (SD-OCT) system was built (Fig. S3). The custom SD-OCT system employed a superluminescent diode source (MT-850-HP, Superlum Inc.) with a 3-dB bandwidth of 160 nm and a central wavelength of about 842 nm. A 50:50 fiber coupler (TW850R5A2, Thorlabs Inc.) was adopted to split the light into the reference and sample arms. The light beam in the reference arm was collimated using a reflective collimator (RC02APC-P01, Thorlabs Inc.). A pair of N-SF11 prisms (Edmund Inc.) were inserted between the reflective collimator and the reflective mirror (PF05-03-P01, Thorlabs Inc.) to compensate for the dispersion imbalance between the reference and sample arms. A homemade rotary joint was mounted on a linear translational stage (X-LSM150A, Zaber Technologies) to rotate and pull back the imaging probe and thus obtain three-dimensional volumetric images. The OCT light back-reflected from reference and sample arms interferences with each other and is collected using a spectrometer (Cobra-S 800, Wasatch Photonics Inc.) with a high line-scan rate up to 250 kHz. To manage the polarization mode dispersion (PMD), two polarization controllers (FPC030, Thorlabs Inc.) are utilized. The imaging depth of the system is about 1.04 mm.

Supplementary Note 4: Choice of optical liquids for microlens fabrication.

Six representative optical liquids, including NOA 60, NOA 61, NOA 63, NOA 65, NOA 68, and NOA 81 (Norland Products Inc.), have been considered to fabricate the microlens (see Table 3). NOA 81 is selected because of (1) fast curing property (20 minutes UV light exposure at a power density of 65 mW/cm² for complete polymerization); (2) good endurance against temperature variation (withstand -150~125°C after polymerization); (3) its low viscosity, which makes it easy to dispense; (4) relatively high elastic modulus, which makes it tough and not easy to deform; and (5) relatively high transmission near 800 nm (as high as 96%).

Table 3: Comparison between optical liquids.

Optical liquid	Cure Time [a]	Working Temperature [b]	Viscosity @ 25 °C [c]	Elastic Modulus [d]	Transmission @ VIS to NIR [e]
NOA60	25 minutes	-15~90°C	300 cps	135000 psi	96%
NOA61	25 minutes	-150~125°C	300 cps	150000 psi	96%
NOA63	40 minutes	-15~90°C	2000 cps	240000 psi	98%
NOA65	40 minutes	-15~60°C	1200 cps	20000 psi	98%
NOA68	40 minutes	-80~90°C	5000 cps	20000 psi	98%
NOA81	20 minutes	-150~125°C	300 cps	200000 psi	96%

[a] – [e] The data is obtained from Norland Product Inc.

Figure S1. Procedures for fabricating liquid-shaped microlens and microendoscope.




Fig. S1 Procedures for fabricating liquid-shaped microlens and microendoscope. **a** The flow chart of microlens fabrication. **b** The procedures for fabricating the liquid-shaped microlens. Different colors of the glass substrate indicate the modification process of the glass surface property: yellow indicates that the hydroxy (-OH) was exposed on the glass surface during oxygen plasma cleaning; green indicates that the fluoride was reacting with the hydroxy; purple indicates that the hydroxy was bonded entirely with the fluoride and the glass substrate was fully fluorinated. **c** The flow chart of liquid-shaped microendoscope fabrication. **d** Corresponding photographs of the fabrication procedures listed in **c**. **e** Simultaneous fabrication of the imaging probes. All scale bars are 200 μm .

Figure S2. Imaging results of simultaneously fabricated microendoscopes.




Fig. S2 Imaging results of simultaneously fabricated microendoscopes. From left to right, the first two images (**a** and **b**) show 2D cross-sectional images of mouse brain obtained from # 1 and # 2 rigid microendoscopes, respectively. The last three images (**c**, **d**, and **e**) show the cross-sectional images of rat esophagus obtained from # 3, # 4, and # 5 flexible microendoscopes, respectively. The imaging results demonstrate the comparable performance of the simultaneously fabricated microendoscopes. # 1 and # 3 are the reported rigid and flexible microendoscopes, respectively. All scale bars are 250 μm .

Figure S3. Schematic of the endoscopic spectral-domain OCT (SD-OCT) system working near 800 nm.




Fig. S3 Schematic of the endoscopic spectral-domain OCT (SD-OCT) system working near 800 nm. SLD: superluminescent diode, PC: polarization controller, FC: fiber coupler, RC: reflective collimator, PP: prism pairs, RM: reflective mirror, TS: translational stage, RJ: rotary joint.

Figure S4. Illustration of OCT scanning in mouse deep brain and the *en face* projection procedures.




Fig. S4 Illustration of OCT scanning in mouse deep brain and the *en face* projection procedures.

Figure S5. Long-term stability on imaging performance of fabricated microendoscopes.




Fig. S5 Long-term stability on imaging performance of fabricated microendoscopes. The 2D cross-sectional images of the rat esophagus, mouse aorta, and mouse deep brain (from top to bottom) obtained when the microendoscopes were fabricated at 1 month (**a**), 3 months (**b**), 6 months (**c**), and 12 months (**d**). The comparable imaging results demonstrate the long-term stability of our microendoscopes. All scale bars are 250 μm .

Figure S6. Calibration of dispensed liquid volume.




Fig. S6 Calibration of dispensed liquid volume. A series of liquid volumes claimed by the dispenser was dispensed on the glass substrate to form the liquid lens, and the volume was measured (with a measurement accuracy of ± 0.05 nL). The slope of the fitted line is about 0.988 and is used to calibrate the claimed dispensed volume.

Figure S7. Refractive index profile of NOA 81 under room temperature.




Fig. S7 Refractive index profile of NOA 81 under room temperature.

Figure S8. Schematic of mouse brain handling for histology.




Fig. S8 Schematic of mouse brain handling for histology. The red dots indicate the burr-hole positions where the glass capillary tubes were inserted. The blue dashed line indicates the transverse direction for brain tissue dissection (the cross-section parallel to the x - z plane).

Supplementary Table 1: Comparison of fabrication methods for OCT microendoscopes.

Methods	Pros	Cons
Liquid shaping-based method	<ol style="list-style-type: none"> 1. Enable monolithic design for minimization of OCT probe. 2. Freeform lens of high design freedom and rotational symmetric surface for aberration corrections. 3. Mass production of freeform lens. 4. Freeform lens of sub-nanometer surface roughness. 5. No need for lens polishing and precise optical alignment. 6. Scalable fabrication. 7. Short endoscope fabrication time of at most 1.5 hours. 8. Fabrication material cost of about \$70, including single-mode fiber, non-core fiber, optical liquid, glass substrate, wetting materials, toque coil, and protective sheath. 	<ol style="list-style-type: none"> 1. Need a relatively expensive liquid dispensing system of about \$30,000 or a home-made liquid dispensing system of about \$6,000.
Two-photon 3D microprinting-based method	<ol style="list-style-type: none"> 1. Enable monolithic design for minimization of OCT probe. 2. No need for lens polishing and precise optical alignment. 3. Freeform lens of high design freedom and complex asymmetric surface for aberration corrections. 	<ol style="list-style-type: none"> 1. Suboptimal surface roughness. 2. Lack of scalability potential. 3. Long printing time up to several hours, depending on the lens size. 4. Expensive two-photon printing machine, ownership cost of at least \$500,000. 5. Limited choice of photo resins.
GRIN fiber-based method	<p>Enable monolithic design for minimization of OCT probe and GRIN fibers are commercially available.</p>	<ol style="list-style-type: none"> 1. Limited choice of GRIN fibers and generally unknown fiber parameters. 2. Lack of capability to correct imaging aberrations. 3. GRIN fibers usually have strong chromatic and spherical aberrations at short wavelength regimes, such as 800 nm. 4. Requires costly and time-consuming angle-polishing. 5. Suboptimal surface roughness due to polishing. 6. Endoscope fabrication time up to several hours.
Fiber ball-lens-based method	<ol style="list-style-type: none"> 1. Enable monolithic design for minimization of OCT probe 2. Provide achromatic performance at short wavelength ranges, such as 800 nm. 	<ol style="list-style-type: none"> 1. Insufficient design freedom and controllability on ball-lens using fiber melting technique. 2. Requires costly and time-consuming angle-polishing. 3. Suboptimal surface roughness due to polishing. 4. Limited to fabricate achromatic endoscopes of less than 1 mm in diameter. 5. Endoscope fabrication time up to several hours.

Supplementary Table 2: Six groups of glass substrates with different parameters of surface wettability modification.

Glass substrate group	Processing time [a]	Volume of fluoride [b]	Contact angle
Glass substrate 1	0 hour	0 μL	$\sim 50^\circ$
Glass substrate 2	4 hours	10 μL	$\sim 67^\circ$
Glass substrate 3	6 hours	10 μL	$\sim 75^\circ$
Glass substrate 4	8 hours	10 μL	$\sim 85^\circ$
Glass substrate 5	12 hours	10 μL	$\sim 90^\circ$
Glass substrate 6	16 hours	20 μL	$\sim 110^\circ$

[a] The processing time indicates the period when the sample is soaked in the fluoride solution.

[b] The fluoride is 1H,1H,2H,2H-Perfluorooctyltriethoxysilane used for modifying the wettability of the glass substrate.

Supplementary Table 3: Two groups of cylinder substrates with different parameters of surface wettability modification.

Cylinder substrate group	Processing time [a]	Volume of fluoride [b]	Contact angle [c]
Circular cylinder substrate	4 hours	10 μL	$\sim 65^\circ$
Elliptical cylinder substrate	8 hours	10 μL	$\sim 80^\circ$

[a] The processing time indicates the period when the sample is baked in a thermotank for fluorination.

[b] The fluoride is 1H,1H,2H,2H-Perfluorooctyltriethoxysilane used for modifying the wettability of the cylinder substrate.

[c] The contact angle is measured when the microlens is not constrained by the physical boundary.

Movie S1: Microlens on a circular cylinder substrate.

Movie S2: Microlens on an elliptical cylinder substrate.