

Angiogenesis biomarkers detection on a polymeric 3D printed device

Original

Angiogenesis biomarkers detection on a polymeric 3D printed device / Palmara, Gianluca; Chiado', Alessandro; Chiappone, Annalisa; Pirri, Candido; Roppolo, Ignazio; Frascella, Francesca. - In: BIOMEDICAL SCIENCE AND ENGINEERING. - ISSN 2531-9892. - ELETTRONICO. - 5:(2021), pp. 93-94. (Intervento presentato al convegno Centro 3R - 3rd Annual Meeting tenutosi a Torino nel September 30th - October 1st) [10.4081/bse.2021.186].

Availability:

This version is available at: 11583/2961478 since: 2022-04-17T10:54:04Z

Publisher:

PAGEPress

Published

DOI:10.4081/bse.2021.186

Terms of use:

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

(Article begins on next page)

Angiogenesis biomarkers detection on a polymeric 3D printed device

Gianluca Palmara,¹ Alessandro Chiadò,¹ Annalisa Chiappone,¹ Fabrizio Pirri,^{1,2} Ignazio Roppolo,¹ Francesca Frascella¹

¹Department of Applied Science and Technology, Politecnico di Torino, Turin;

²Sustainable Future Technologies @Polito, Istituto Italiano di Tecnologia, Turin, Italy

Abstract

Polymeric 3D printed chips with intrinsic tuned functionalities were produced and employed for the detection of angiogenesis biomarkers.

Introduction

The monitoring of tumoral biomarkers involved in pathological angiogenesis is an important task in the establishment of an early diagnostics for patients. Biosensors could represent a promising platform for this purpose. These devices require complex chemical functionalizations to immobilize the biorecognition elements. Here, a novel biosensor consisting in a polymeric 3D-printed chip is presented. The device is printed in a single-step procedure with tuned chemical functionalities, later exploited for the tethering of antibodies. The detection of Vascular Endothelial Growth Factor (VEGF) and Angiopoietin-2 (Ang-2), well-known cancer biomarkers, was performed in order to prove its reliability as a screening tool.

Materials and Methods

The polymeric chips were realized using a Digital Light Processing 3D printer.¹ The revealment of the biomarkers of interest and the estimation of their Limit of Detection (LOD) and Limit of Quantification (LOQ) were performed by means of a sandwich ELISA-like immunoassay.

Results

Acrylic acid was added to three resins, to introduce carboxyl groups in the polymeric matrix, to be exploited as anchor point for the immobilization of any bioreceptor that exposes a primary amine.

The printed polymers were tested for their physicochemical properties and protein grafting capabilities, in order to select the top performing formulation that was afterward used for the production of two Lab-on-Chip devices.

The first chip consisted of a microfluidics responsible to convey samples and buffers to an incubation well, where the specific biomolecular interactions took place. The second chip was an alternative version, with three incubation wells simply accessible through an inlet port (see Figure 1).

The LOD evaluated for VEGF and Ang-2 were compared with the cut-off values that are reported in the literature.^{2,3} VEGF sensitivity was not high enough to reveal abnormal levels of this growth factor. On the other hand, the LOD relative to Ang-2 (0.8 ng mL⁻¹) was below the threshold concentration (2.5 ng mL⁻¹), demonstrating a good reliability in a suitable range of values (see Figure 2).

Correspondence: Gianluca Palmara, Department of Applied Science and Technology, Politecnico di Torino, C.so Duca degli Abruzzi 24, 10129Torino, Italy. Tel.: +39.0110907329 E-mail: gianluca.palmara@polito.it

Key words: Biochemical functionalization; biosensors; 3D-printing; ELISA.

Acknowledgments: This work was performed in the framework and financed by POLITO BIOMed LAB, an interdepartmental laboratory financed by Politecnico di Torino, DEFLeCT project funded by Regione Piemonte and FOOD DRUG FREE project (funded by Regione Piemonte).

Disclosures : There are no conflicts to declare.

Conference presentation: This paper was presented at the Third Centro 3R Annual Meeting - L'era delle 3R: modelli *in silico*, *in vitro* e *in vivo* per promuovere la ricerca traslazionale - 30 September - 1 October 2021, Evento online organizzato dal Politecnico di Torino.

Received for publication: 9 July 2021.

Accepted for publication: 7 September 2021.

This work is licensed under a Creative Commons Attribution NonCommercial 4.0 License (CC BY-NC 4.0).

©Copyright: the Author(s), 2021

Licensee PAGEPress, Italy

Biomedical Science and Engineering 2021; 4(s1):186
doi:10.4081/bse.2021.186

Discussion and Conclusions

The combination of the 3D-printing technology with an improvement of photocurable formulations allowed to design a polymeric device having intrinsic tuned




Figure 1. A) CAD model of the LOC; B) microscope images of the 3D printed microfluidic device; C) images of the printed modular chip.




Figure 2. Detection of Ang-2 through sandwich ELISA-like protocol in the modular LOC: the graph shows the OD signals obtained for different concentrations of Ang-2.

functionalities with no need for further chemical derivatization. This system was able to specifically recognize Ang-2 with a LOD compatible with the range of concentrations that are required for the identification of a malignant scenario.

References

1. Chiadò A, Palmara G, Chiappone A, et al. A modular 3D printed lab-on-a-chip for early cancer detection. *Lab Chip* 2020;20:665-74.
2. Monastero RN, Pentyla S. Cytokines as biomarkers and their respective clinical

cutoff levels. *Int J Inflamm* 2017;2017:4309485.

3. Li P, He Q, Luo C, Qian L. Diagnostic and prognostic potential of serum angiopoietin-2 expression in human breast cancer. *Int J Clin Exp Pathol* 2015;8:660-4.

Non commercial use only